

Product Name :
Electrosurgical Unit Max Output Power : 30w

Product Code :
OT Equipment0014



Description :

Electrosurgical Unit Max Output Power : 30w Characteristic : Control the power by using the power adjustment button of the ESU pencil so as to control output power and the operation time is shortened. Dual microprocessors (one is to control function & the second is to terminate activation if a hazard or safety issued is detected. Memory Function : Set up the requested function parameter after turn-off. At the next startup the previous setting parameter will be displayed. From 10W to 30W, the output of each watts can be manipulated precisely. And below 10W, the adjustment is done at the 0.1W amplitude by up and down, all of micro surgery is effected by this machine. Stimulate the argon system for activating the spray coagulation Output Mode : Monopolar Coagulation : (30W, Load : 1000?) Bipolar Coagulation 1 : (30W, Load : 200?) Scope of Application : Dermatology, Plastic Surgery, Proctology, Gynecology, Ophthalmology, Urology, Otolaryngology, Dental / Oral Surger Standard Matching Accessories : Operating ambient temperature : 5°C ~ 40°C Relative humidity range:# 80% Pressure range : 86.0 ~ 106.0 kPa Power Supply:220 / 110 V, 50/60 Hz Operation Frequency : 450 kHz \pm 50kHz Transportation Data : Size : 22.5cm (L) *19.5cm (W) *12cm (H) (ESU UNIT size) ; 33 (L) *25 (W) *26.5cm (H) (outer carton) Weight : 2.5kg (Net Weight) ; 4kg (Gross Weight) Standard Matching Accessories : Disposable power control pencil (optional) : 2 pcs Or reusable power control pencil (optional) : 1 pc Disposable solid electrosurgical pad and reusable cable (optional) : 1 pc Or reusable stainless steel pad and cable (optional) Power Cord : 1pc Reusable Bipolar Forceps and Cable : 1 pc Individual Pedal Switch : 1 pc Instruction Manual : 1 pc

Website: SITE,

NAUGRA[®]

Medical Hospital Equipments Manufacturers